


50X1-HUM



M E T E R I U

TECHNICAL DESCRIPTION

Zh 3.360.031 TO

S E C R E T



50X1-HUM

50X1-HUM




METER NY

TECHNICAL

DESCRIPTION

W3.360.031-TO

SECRET



50X1-HUM

50X1-HUM

CONTENTS

| | Page |
|---|-------|
| I. TECHNICAL DATA | 3 |
| 1. Purpose | 3 |
| 2. Technical Data | 3 |
| II. COMPONENTS | 4 |
| III. PRINCIPLE OF OPERATION | 4 |
| 3. Block-Diagram of Meter | 4 |
| 4. Description of Key Circuit Diagram ... | 5 |
| 5. Design | 8 |
| IV. MAINTENANCE AND OPERATION | 8 |
| 6. Preparation for Operation | 8 |
| 7. Operating Instructions | 10 |
| 8. Switching Off the Meter | 11 |
| 9. Troubleshooting | 11 |
| 10. Valve Chart | 13 |
| 11. D.C. Operating Conditions | 13 |
| 12. Valve Basing | 14 |
| 13. Table of Winding Data | 15 |
| <u>Appendix:</u> Key Circuit Diagram #6.294.006Cx3..... | Inset |
| and Specifications | 17 |

50X1-HUM

SECRET

50X1-HUM

I. TECHNICAL DATA

1. PURPOSE

The MY instrument is a level meter designed for measuring the root-mean-square value of sinusoidal, noise and D.C. voltages. The instrument is used to check receiving and amplifying channels, as well as to measure the interference voltage and perform other measurements.

2. TECHNICAL DATA

(a) Meter MY has the following measurement ranges:

A.C. voltage:

0.03-0.1-0.3-1-3-10-30-100 V;

D.C. voltage:

1-3-10-30-100-300-1000 V.

(b) The frequency range of the meter is 40 c.p.s. to 100 Kc/s.

(c) Accuracy of sinusoidal and noise voltage measurements is within 10 mV to 100 V $\pm 5\%$.

Accuracy of D.C. voltage measurement ranges from 0.1 to 1000 V not below $\pm 5\%$.

(d) Input impedance of the A.C. voltmeter is 0.5 megohm, and that of the D.C. voltmeter, 10 megohms.

Input capacitance of the A.C. voltmeter is 15 pF.

(e) The voltage of 10 mV is considered as a zero level of the decibel scale.

(f) The instrument operates normally at an ambient temperature of -5°C to $+40^{\circ}\text{C}$, at a relative humidity up to 95%, and after a prolonged stay at a temperature below -40°C with the subsequent stay of the instrument at a temperature of $+20^{\circ}\text{C} \pm 5^{\circ}\text{C}$.

3

SECRET

50X1-HUM

50X1-HUM

- (g) The instrument can operate continuously for 8 hours.
 (h) The instrument is supplied with 127 or 220 V $\pm 10\%$, 50 c.p.s. $\pm 3\%$ by the A.C. mains. The consumed power does not exceed 35 VA.
 (i) The instrument weights not more than 10 kg.
 (j) The instrument dimensions are 314x190x198 mm.

II. COMPONENTS

The instrument includes the following components:

- (a) meter IV;
- (b) supply cable;
- (c) two measuring cables;
- (d) canvas cover.

III. PRINCIPLE OF OPERATION

3. BLOCK-DIAGRAM OF METER

The level-meter block-diagram is shown in Fig. 1.

With switch Π set to positions 1 or 2, the level meter functions as an A.C. electronic voltmeter.




FIG. 1. BLOCK DIAGRAM OF METER IV

4

SECRET

50X1-HUM

50X1-HUM

With switch Π in position 1, the meter measures the root-mean-square value of sinusoidal voltage, and in position 2, the effective value of noise voltage. Via a voltage divider the A.C. voltage is applied to the wide-band amplifier and then to the detector. From the detector output the rectified voltage is fed to the indicator. With switch Π in positions 3 or 4, the meter functions as a D.C. voltmeter.

With switch Π in position 3, the meter measures positive D.C. voltage relative to the "earth", whereas with switch Π in position 4, the meter measures negative D.C. voltage relative to the "earth".

A D.C. source voltage is applied to the D.C. voltage divider and then to the measuring bridge, into the diagonal of which an indicator is cut in.

The circuit is provided with a limiter which protects the indicator from overloading in case the measuring range on the divider is set wrong.

Both the A.C. and D.C. voltages are read off with the aid of the same indicator scales.

4. DESCRIPTION OF KEY CIRCUIT DIAGRAM

(See Diagram $\times 6.294.006 \times 3$)

(A) A.C. Voltmeter

The A.C. voltmeter employs valves 6H1 Π (13), 6X2 Π (21, 31) and 6X2 Π (38).

There is a stage with a cathode load (valve 13) at the electronic voltmeter input. A low input capacitance of the stage ensures a constant high input resistance up to the frequency of 100 Kc/s.

Both triodes of valve 6H1 Π (13) are connected in parallel in order to decrease its inner resistance, which is required to ensure linearity of the stage amplitude characteristic within 100 V. This excludes a necessity to use a high-ohmic divider at the voltmeter input which might lead to a considerable increase of input capacitance and, consequently, to the narrowing of the frequency range. The voltage divider resistors (5 to 12) serve as a stage cathode load.

5

SECRET

50X1-HUM

50X1-HUM

Switch TRANSFORMER TAP (ПОДКЛЮЧЕНИЕ НАПРЯЖЕНИЯ) 15 allows to select one of the required ranges.

The circuit employs a low-ohmic voltage divider (30 kilohms) which makes its frequency characteristic almost independent of the input capacitance of the subsequent stage.

The wide-band voltage amplifier employs valves 6X2П (21 and 31) in accordance with the ordinary rheostat circuit with a negative feedback dependent on values of resistors 23, 24, 37 and capacitor 33.

The feedback provides for the linearity of the frequency characteristic of the amplifier and stability of its operation.

Capacitor 33 is designed for compensating the frequency characteristic at the high frequencies. Potentiometer 40 is intended to set the electronic-voltmeter zero.

To make it possible to measure the sinusoidal and noise voltages at one scale, potentiometer 80 is connected in parallel with the indicator for measuring the sinusoidal voltage. To make the indicator time constant equal to 4 to 6 sec., which is required for measuring the noise voltage, charging capacitor 78 is used. When the D.C. voltage is measured, capacitor 78 is discharged through resistor 77. Capacitor 75 serves to provide for required inertia of the indicator pointer in the process of measuring the sinusoidal voltages.

(B) D.C. Voltmeter

The D.C. voltmeter circuit is a bridge consisting of four arms. Instrument M-24, serving as an electronic-voltmeter indicator, is cut into the bridge diagonal.

Two arms of the bridge are triodes of valve 6H1П (68), whereas the third and fourth arms are resistors 70 and 71 interconnected by potentiometer 69 which is designed for setting the electronic-voltmeter zero.

Diode 64 serves to limit the current flowing through the instrument when a voltage exceeding 4 V is applied to the left-hand triode grid of valve 68. The limiting proceeds in the following way: the anodes of valve 64 are supplied with

6

SECRET

50X1-HUM

50X1-HUM

-3.8 V relative to the cathodes, and the valve is cut off. When the voltage at the grid of valve 68 exceeds the value ± 3.8 V (depending on the polarity of the voltage to be measured, the left-hand or right-hand diode is operative), the current begins to flow through the diode. As a result, the voltage drops across resistor 63 which prevents further increase of the voltage across the grid of valve 68.

A voltage divider employing resistors 56 to 62 is connected to the voltmeter input.

The total resistance of the divider is equal to 10 megohms. Resistor 63 decreases the influence of the variation of the grid-leak resistor of valve 68 at different positions of switch MEASUREMENT RANGES 15, thus nearly excluding the necessity to compensate the instrument zero when switching over from one scale to another.

Besides, resistor 63 and capacitor 66 serve as a filter protecting the grid circuit of valve 68 from possible variable components of the voltage when measuring the D.C. voltages, as it may take place, for instance, in the process of measuring the anode voltage of the operative amplifier.

Potentiometer 67 is designed to adjust the circuit sensitivity.

The measurement ranges of both voltmeters are switched over with the aid of switch 15. Selector switch KIND OF MEASUREMENT (РОД ИЗМЕРЕНИЙ) 76 serves to select the required kind of measurements.

The circuit is supplied by the A.C. mains of 220 or 127 V, which is achieved by switching over correspondingly power-transformer 49 primary winding with the aid of fuses 51 and 54.

The rectifier employs a full-wave circuit with valve 48. A filter composed of choke 45 and capacitors 44, 46, 47 serves to smooth pulsations.

The A.C. voltmeter supply voltage is not stabilized, since a sufficient stability of the circuit is ensured at the expense of a deep negative feedback.

7

SECRET

50X1-HUM

50X1-HUM

5. DESIGN

(See Figs 2 and 3)

All elements of the meter are mounted on the chassis consisting of two main plates: a vertical plate and a horizontal one fastened together. At the same time, the vertical plate serves as a front panel of the meter and carries all controls. There are two handles on the panel which are used to remove the chassis out of the housing and to protect simultaneously the controls from damage when the meter is placed with its front panel downwards for repair.

The meter chassis carries all elements of the circuit.

The indicator is arranged on the meter front panel and is made shockproof by means of a rubber gasket. All transformers, chokes and wire-wound resistors are impregnated with special lacquer to prevent moisture from penetrating inside the windings. All soldered joints are painted for the purpose of checking. The chassis is inserted into the housing and fastened by four screws. For the purpose of conveniently carrying the meter, its casing is made in the form of a suitcase with a detachable lid covering the front panel. The inside of the meter lid is provided with clamps which serve for arrangement of a supply cable and two measuring cables. On the rear side of the chassis, under the flap, are mounted: a supply block and two fuses for 127 and 220 V. This construction allows to change over the mains voltages and to replace the fuses without removing the meter casing. The walls of the casing are provided with louvers for better cooling of internal elements. The meter is provided with a cover to prevent the former from damage and penetration of moisture during transportation.

IV. MAINTENANCE AND OPERATION

6. PREPARATION FOR OPERATION

To prepare the meter for operation, proceed as follows:

- (a) take the meter out of the canvas cover, unfasten the locks and remove the lid from the hinges;

SECRET

50X1-HUM

50X1-HUM




FIG. 2. FRONT VIEW OF THE METER




FIG. 3. REAR VIEW OF THE METER (WITH CASING REMOVED)

SECRET

50X1-HUM

50X1-HUM

(b) push aside the flap located on the rear wall of the meter. Check whether the fuse is inserted correctly. The fuse should be inserted into the socket, the inscription of which corresponds to the voltage value of the supply mains (127 or 220 V). Couple the supply cable connector to the voltage source;

(c) connect reliably the meter terminal marked $\frac{1}{\equiv}$ with the body of the object under test with a flexible copper wire with a cross-section of at least 2.5 mm²;

(d) set toggle switch MAINS (CETb) to position OFF (BKU.);

(e) couple the supply cable to the mains and set toggle switch MAINS to position ON (BKU.). While doing so, the signal lamp will light up. After that the meter is ready for operation.

7. OPERATING INSTRUCTIONS

Warning

Do not connect the meter to the A.C. mains if it is not in the casing.

Mind that some parts of the meter are energized with a voltage dangerous to life.

Measurement of Noise and Sinusoidal Voltages

(a) Set selector switch KIND OF MEASUREMENT to position A.C. VOLTAGE (V~);

(b) take the measuring cables with lugs out of the lid and plug them into terminals marked $\frac{1}{\equiv}$ and A.C. VOLTAGE (V~);

(c) interconnect terminals $\frac{1}{\equiv}$ and A.C. VOLTAGE (V~) by a short rigid jumper and, operating knob A.C. 0 SETTING (YCTAHOBIA 0~), set the meter pointer to zero;

(d) set selector switch MEASUREMENT RANGES to position 100 V;

(e) using the cables, connect the voltage to be measured, select the required measurement range with the aid of selector switch MEASUREMENT RANGES and read the voltage off the scale.

10

SECRET

50X1-HUM

50X1-HUM

When measuring the sinusoidal voltage, selector switch KIND OF MEASUREMENT should be in position A.C. (\sim), and when measuring the noise voltage, in position NOISE (ШУМ).

Measurement of D.C. Voltage

- (a) Set selector switch KIND OF MEASUREMENT to position D.C. VOLTAGE (V=);
- (b) take the measuring cables with lugs out of the lid and plug them into terminals \perp and D.C. VOLTAGE (V=);
- (c) using the cables, interconnect terminals \perp and D.C. VOLTAGE (V=) and, operating knob D.C. 0 SETTING (УСТАНОВКА 0=), set the meter pointer to zero;
- (d) select the required measurement range with the aid of selector switch MEASUREMENT RANGES;
- (e) connect the voltage to be measured so that terminal \perp is always coupled with the grounded pole of the voltage under measurement. In case the meter pointer deflects to the left, set selector switch KIND OF MEASUREMENT to the sign of the opposite polarity;
- (f) read the measured voltage off the indicator scale.

8. SWITCHING OFF THE METER

- (a) Set toggle switch MAINS to position OFF;
- (b) disconnect the supply cable from the mains;
- (c) remove the supply cable from the socket and close the flap;
- (d) disconnect the grounded wire from the ground terminal of the meter;
- (e) place the supply cable into the lid. Set the latter onto the hinges and fasten the locks;
- (f) place the meter in the canvas cover.

9. TROUBLESHOOTING

If any troubles occur in the meter, they should be eliminated by persons well acquainted with the circuit diagram and the principle of operation of the meter. In case some

11

SECRET

50X1-HUM

50X1-HUM

troubles take place which are not listed below, the meter should be repaired in a workshop by skilled personnel.

A list of possible troubles which can be eliminated in operation is given below.

| Trouble | Cause | Remedy |
|--|--|---|
| 1 | 2 | 3 |
| 1. Meter switched on, signal lamp fails to light up | (a) Burnt out fuse (b) Burnt out signal lamp (c) No mains voltage (d) Broken supply cable | (a) Replace fuse (b) Replace lamp (c) Check for presence of voltage in mains (d) Check cable and eliminate break |
| 2. Meter fails to operate in all modes of operation, signal lamp does not light | (a) Faulty valve 48 (b) Broken one of electrical capacitors 44 or 47 | (a) Replace valve, using set of spares (b) Send meter to workshop for repair |
| 3. A.C. voltmeter fails to operate | Faulty one of valves 21, 31 or 38 | Replace faulty valve, using set of spares |
| 4. Potentiometer 0 SETTING (УСТАНОВКА 0) fails to operate. Indicator pointer deflects to left beyond scale | Faulty valve 38 | Replace valve, using set of spares |

50X1-HUM

50X1-HUM

| 1 | 2 | 3 |
|---|-----------------|------------------------------------|
| 5. Zero cannot be set by potentiometer O SETTING when measuring D.C. voltage | Faulty valve 68 | Replace valve, using set of spares |

The valve replacement: When the valves are replaced, the meter does not require any additional adjustment.

10. VALVE CHART

| No. | Valve No. in diagram | Valve type |
|-----|----------------------|------------|
| 1 | 13 | 6H1Π |
| 2 | 21 | 6X2Π |
| 3 | 31 | 6X2Π |
| 4 | 38 | 6X2Π |
| 5 | 48 | 6U4Π |
| 6 | 50 | MR-14 |
| 7 | 64 | 6X2Π |
| 8 | 68 | 6H1Π |

11. D.C. OPERATING CONDITIONS

| No. | Valve No. in diagram | Valve type | Valve operating conditions | | | | | | | | | | Note |
|-----|----------------------------|---------------|----------------------------|-----|-----|---|-----|-----|-----|-----|---|--|------|
| | | | Electrode No. | | | | | | | | | | |
| | | | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | | |
| 1 | 13 | 6H1Π | 300 | - | 150 | - | - | 300 | - | 150 | - | | |
| 2 | 21 | 6X2Π | - | 2.0 | - | - | 140 | 110 | 2.0 | - | - | | |
| 3 | 31 | 6X2Π | - | 2.3 | - | - | 200 | 100 | 2.3 | - | - | | |
| 4 | 64 | 6X2Π | 2.2 | 5.0 | - | - | - | - | - | - | - | | |
| 5 | 68 | 6H1Π | 70 | - | 1.4 | - | - | 70 | - | 1.4 | - | | |

The voltage between the connection point of resistors 14, 4, 5 and the cathode of valve 13 is 3 V.

13


SECRET

50X1-HUM

50X1-HUM

All voltages are measured relative to the earth.
 Permissible deviations from the given values are ± 10 per cent.

12. VALVE BASING





SECRET

50X1-HUM

50X1-HUM

13. TABLE OF WINDING DATA


| Ref. No. in diagram | Description | Electrical and other data |
|---------------------------|--|---|
| 49 | <p>Power transformer Ж 6.722.003</p>  | <p>Ist winding ПЗЛ, dia.0.41, 1-2=915 turns IIInd winding ПЗЛ, dia.0.1, 4-5=5-6=2115 turns IIIrd winding ПЗЛ, dia.0.59, 7-8=50 turns IVth winding ПЗЛ, dia.1.08, 9-10=50 turns Magnetic circuit 34AA III20x30</p> |
| 46 | <p>Filter choke Ж 6.766.020</p>  | <p>ПЗЛ, dia.0.12, 12,500 turns Magnetic circuit 34AA III12x12</p> |

15

SECRET

50X1-HUM


50X1-HUM



SECRET

50X1-HUM


50X1-HUM



SECRET

50X1-HUM


50X1-HUM



SECRET

50X1-HUM

50X1-HUM



SECRET

50X1-HUM



50X1-HUM

50X1-HUM

Specifications for Key Circuit Diagram

| Ref. No. | St. Std, Specifications, Dwg | Description | Rating | Qty | Note | Modification |
|----------|------------------------------|-------------------------------|-------------|----------------|---|--------------|
| 1 | 2 | 3 | 4 | 5 | 6 | 7 |
| 1 | Y7.527.014 | Jack | | 1 | | |
| 2 | FOCT 5561-54 | Capacitor | 20 μ F | 1 | | |
| 3 | FOCT 6119-54 | K3-16-400-20M Capacitor | 5600 pF | 1 | | |
| 4 | OMO.467.003TV | KCO-5-500-B-5600-I Resistor | 0.51 megohm | 1 | | |
| 5 | OMO.467.003TV | MNT-1-0.51-I-A Resistor | 18 kilohms | 1 | | |
| | OMO.467.003TV | MNT-1-18,000-I-A Resistor | 1600 ohms | One of ratings | Selected in tuning, series-connected, 20,500 ohms $\pm 1\%$ | |
| | OMO.467.003TV | MNT-1-1600-I Resistor | 2000 ohms | | | |
| | OMO.467.003TV | MNT-1-2000-I Resistor | 2400 ohms | | | |
| | OMO.467.003TV | MNT-1-2400-I Resistor | 2700 ohms | | | |
| | OMO.467.003TV | MNT-1-2700-I Resistor | 3000 ohms | | | |
| | OMO.467.003TV | MNT-1-3000-I Resistor | 3300 ohms | | | |
| | OMO.467.003TV | MNT-1-3300-I Resistor | 5600 ohms | | | |
| 6 | OMO.467.003TV | MNT-1-5600-I Resistor | 620 ohms | One of ratings | Same, 6500 ohms $\pm 1\%$ | |
| | OMO.467.003TV | MNT-1-620-I Resistor | 1200 ohms | | | |
| | OMO.467.003TV | MNT-1-1200-I Resistor | 1800 ohms | | | |
| 7 | OMO.467.003TV | MNT-1-1800-I Resistor | 160 ohms | One of ratings | Same, 2050 ohms $\pm 1\%$ | |
| | OMO.467.003TV | MNT-1-160-I Resistor | 200 ohms | | | |
| | OMO.467.003TV | MNT-1-200-I Resistor | 240 ohms; | | | |
| | OMO.467.003TV | MNT-1-240-I; MNT-1-270 ohms-I | 270 ohms | | | |
| | OMO.467.003TV | MNT-1-330-I Resistor | 300 ohms | | | |
| | OMO.467.003TV | MNT-1-300-I Resistor | 560 ohms | One of ratings | Same, 650 ohms $\pm 1\%$ | |
| 8 | OMO.467.003TV | MNT-1-560-I Resistor | 100 ohms | | | |
| | TYOZO.467.004 | BC-0.25-100-I Resistor | 120 ohms | | | |
| | OMO.467.003TV | MNT-1-120-I Resistor | | | | |

17

SECRET

50X1-HUM

50X1-HUM

| 1 | 2 | 3 | 4 | 5 | 6 | 7 |
|----|---------------|---------------------------------|---------------------------|---|--|---|
| 9 | OMO.467.003TV | Resistor MNT-1-130-I | 130 ohms | 1 | Selected in tuning, One of series- rat- ings 205 ohms ±1% | |
| | TYOMO.467.004 | Resistor BC-0.25-68-I | 68 ohms | | | |
| | TYOMO.467.004 | Resistor BC-0.25-75-I | 75 ohms | | | |
| | TYOMO.467.004 | Resistor BC-0.25-82-I | 82 ohms | | | |
| 10 | X6.732.007 | Wire-wound resis- tor | 650 ohms ±1% | 1 | | |
| 11 | X6.732.005 | Wire-wound resis- tor | 20.5 ohms ±1% | 1 | Same | |
| 12 | X6.732.006 | Wire-wound resistor | 10.2 ohms ±1% | 1 | Same | |
| 13 | 4TV.01.105.53 | Valve 6H1H | | 1 | | |
| 14 | OMO.467.003TV | Resistor MNT-1-680-I | 680 ohms | 1 | | |
| 15 | HO.360.006 | Switch 11THH-KB | 11 positions, 4 wafers | 1 | Porcelain | |
| 16 | OMO.467.003TV | Resistor MNT-1-0.15-I-A | 0.15 megohm | 1 | | |
| 17 | FOCT 6118-52 | Capacitor K3-2-400-20M | 20 μF | 1 | | |
| 18 | FOCT 6118-52 | Capacitor KBT-M-400-0.03-II | 0.03 μF | 1 | | |
| 19 | OMO.467.003TV | Resistor MNT-1-18,000-I-A | 18 kilohms | 1 | | |
| 20 | OMO.467.003TV | Resistor MNT-1-62,000-I-A | 62 kilohms | 1 | | |
| 21 | 4TV.01.104.53 | Valve 6X2H | | 1 | | |
| 22 | OMO.467.003TV | Resistor MNT-1-0.43-I-A | 0.43 megohm | 1 | | |
| 23 | OMO.467.003TV | Resistor MNT-1-390-I | 390 ohms | 1 | | |
| 24 | B6.731.002 | Potentiometer 200 ohms | 200 ohms | 1 | | |
| 25 | FOCT 6119-54 | Capacitor KCO-5-500-B-5100-I | 5100 pF | 1 | | |
| 26 | OMO.462.022TV | Capacitor MBFN-2-400-1-II | 1 μF; 400 V | 1 | | |
| 27 | OMO.467.003TV | Resistor MNT-1-1-I-A | 1 megohm | 1 | | |
| 28 | OMO.467.003TV | Resistor MNT-1-0.2-I-A | 0.2 megohm | 1 | | |
| 29 | OMO.467.003TV | Resistor MNT-1-18,000-I-A | 18 kilohms | 1 | | |
| 30 | OMO.467.003TV | Resistor MNT-1-47,000-I-A | 47 kilohms | 1 | | |

18

SECRET

50X1-HUM

50X1-HUM

| 1 | 2 | 3 | 4 | 5 | 6 | 7 |
|----|----------------|--|----------------------|---|---------------------------------|---|
| 31 | 4TY 01.104-53 | Valve 6X2П | | 1 | | |
| 32 | OKO.467.003TY | Resistor | 820 ohms | 1 | | |
| 33 | OKO.461.015TY | MJT-1-820-I Capacitor | | | | |
| | | KCO-5-500-B-3000-II; KCO-5-250-B-10,000- -II | 3000 to 10,000 pF | 1 | Installed, if neces- sary | |
| 34 | FOCT 5561-54 | Capacitor | 50 μF; 30 V | 1 | | |
| | | K3-16-30-50M | | | | |
| 35 | OKO.462.022TY | Capacitor | 0.25 μF; 400V | 1 | | |
| | | MEPH-2-400-0.25-II | | | | |
| 36 | OKO.462.022TY | Capacitor | 1 μF; 400 V | 1 | | |
| | | MEPH-2-400-1-II | | | | |
| 37 | OKO.467.003TY | Resistor | 82 kilohms | 1 | | |
| | | MJT-1-82,000-I-A | | | | |
| 38 | 4TY 01.108-53 | Valve 6X2П | | 1 | | |
| 39 | OKO.462.022TY | Capacitor | 0.5 μF | 1 | | |
| | | MEPH-2-400-0.5-II | | | | |
| 40 | OKO.468.004TY | Resistor | 100 kilohms | 1 | | |
| | | CH-I-1-100-A-13 | | | | |
| 41 | OKO.467.003TY | Resistor | 15 kilohms | 1 | | |
| | | MJT-1-15,000-I-A | | | | |
| 42 | OKO.468.004TY | Resistor | 3.3 kilohms | 1 | | |
| | | CH-I-1-3.3A ₄ | | | | |
| 43 | OKO.467.003TY | Resistor | 33 kilohms | 1 | | |
| | | MJT-2-33 kilohms-I-A | | | | |
| 44 | FOCT 5561-54 | Capacitor | 20 μF | 1 | | |
| | | K3-2-400-20M | | | | |
| 45 | X6.775.020 | Choke | W12x12 | 1 | | |
| 46 | FOCT 6118-52 | Capacitor | 0.1 μF | 1 | | |
| | | KET-M-200-0.1-II | | | | |
| 47 | FOCT 5561-54 | Capacitor | 10 μF | 1 | | |
| | | K3-2-400-10M | | | | |
| 48 | 4TY 01.109-53 | Valve 6X2П | | 1 | | |
| 49 | X6.722.003 | Transformer | W20x30 | 1 | | |
| 50 | HC40.337.001TY | Valve MH-14 | 63 V; 0.28 A | 1 | | |
| 51 | FOCT 5010-53 | Fuse PK-45-0.5 | 0.5 A | 1 | | |
| 52 | Y7.528.031 | Supply block | | 1 | | |
| 53 | HMO360.606 | Tumbler TB2-1 | 220 V | 1 | | |
| 54 | FOCT 5010-53 | Fuse PK-45-0.25 | 0.25 A | 1 | | |
| 55 | Y7.527.014 | Jack | | 1 | | |
| 56 | OKO.467.003TY | Resistor | 6.2 megohms | 1 | Selected | |
| | | MJT-1-6.2-1-A | | | in | |
| | OKO.467.003TY | Resistor | 360 kilohms | 1 | One | |
| | | MJT-1-0.3-I-A | | | of | |
| | | Resistor MJT-1- | 0.3 megohms | 1 | rat- | |
| | | -360 megohms-I-A | | | ings | |
| | OKO.467.003TY | Resistor | 0.43 megohms | 1 | 6.84 | |
| | | MJT-1-0.43-I-A | | | megohms | |
| | | | | | ±1% | |

19

SECRET

50X1-HUM

50X1-HUM

| 1 | 2 | 3 | 4 | 5 | 6 | 7 |
|----|---------------|--|--------------|--------------------------------|--|---|
| | | Resistor MNT-1-510 kilohms- -I-A | 510 kilohms | One of rat- ings | Selected in tuning, series- connected, 6.84 megohms ±1% | |
| | OMO.467.003TY | Resistor MNT-1-0.56-I-A | 0.56 megohms | | | |
| | | Resistor MNT-1-620 kilohms- -I-A | 620 kilohms | | | |
| | OMO.467.003TY | Resistor MNT-1-0.75-I-A | 0.75 megohm | | | |
| | | Resistor MNT-1-680 kilohms- -I-A | 680 kilohms | | | |
| | OMO.467.003TY | Resistor MNT-1-0.82-I-A | 0.82 megohm | | | |
| | OMO.467.003TY | Resistor MNT-1-0.91-I-A | 0.91 megohm | | | |
| 57 | OMO.467.003TY | Resistor MNT-1-2-I-A | 2 megohms | 1 One of rat- ings | Same, 2.16 megohms ±1% | |
| | OMO.467.003TY | Resistor MNT-1-62,000-I-A | 62 kilohms | | | |
| | | Resistor MNT-1-75 kilohms-I-A | 75 kilohms | | | |
| | OMO.467.003TY | Resistor MNT-1-0.1-I-A | 0.1 megohm | | | |
| | | Resistor MNT-1-19 kilohms-I-A | 19 kilohms | | | |
| | OMO.467.003TY | Resistor MNT-1-0.15-I-A | 0.15 megohm | | | |
| | | Resistor MNT-1-120 kilohms-I-A | 120 kilohms | | | |
| | OMO.467.003TY | Resistor MNT-1-0.2-I-A | 0.2 megohm | | | |
| | | Resistor MNT-1-180 kilohms-I-A | 180 kilohms | | | |
| | OMO.467.003TY | Resistor MNT-1-220 kilohms-I-A | 220 kilohms | | | |
| | | Resistor MNT-1-0.24-I-A | 0.24 kilohms | | | |
| 58 | OMO.467.003TY | Resistor MNT-1-0.62-I-A | 0.62 megohm | 1 One of rat- ings | Same, 0.684 megohm ±1% | |
| | OMO.467.003TY | Resistor MNT-1-30,000-I-A | 30 kilohms | | | |
| | | Resistor MNT-1-36 kilohms-I-A | 36 kilohms | | | |
| | OMO.467.003TY | Resistor MNT-1-43,000-I-A | 43 kilohms | | | |
| | | Resistor MNT-1-51 kilohms-I-A | 51 kilohms | | | |

20

SECRET

50X1-HUM

50X1-HUM

| 1 | 2 | 3 | 4 | 5 | 6 | 7 |
|----|---------------|-----------------------|-------------|------------------|--|---|
| | OXO.467.003TY | Resistor | 56 kilohms | One of ratings | Selected in tuning, series-connected, 0.684 megohm $\pm 1\%$ | |
| | | MJT-1-56,000-I-A | | | | |
| | | Resistor | 62 kilohms | | | |
| | | MJT-1-62 kilohms-I-A | | | | |
| | OXO.467.003TY | Resistor | 75 kilohms | | | |
| | | MJT-1-75,000-I-A | | | | |
| | | Resistor | 68 kilohms | | | |
| | | MJT-1-68 kilohms-I-A | | | | |
| | OXO.467.003TY | Resistor | 82 kilohms | | | |
| | | MJT-1-82,000-I-A | | | | |
| | OXO.467.003TY | Resistor | 91 kilohms | | | |
| | | MJT-1-91,000-I-A | | | | |
| 59 | OXO.467.003TY | Resistor | 0.2 megohm | 1 One of ratings | Same, 0.216 megohm $\pm 1\%$ | |
| | | MJT-1-0.2-I-A | | | | |
| | OXO.467.003TY | Resistor | 6200 ohms | | | |
| | | MJT-1-6200-I | | | | |
| | OXO.467.003TY | Resistor | 7.5 kilohms | | | |
| | | MJT-1-7.5 kilohms-I-A | | | | |
| | OXO.467.003TY | Resistor | 10,000 ohms | | | |
| | | MJT-1-10,000-I-A | | | | |
| | OXO.467.003TY | Resistor | 15 kilohms | | | |
| | | MJT-1-15,000-I-A | | | | |
| | | Resistor | 12 kilohms | | | |
| | | MJT-1-12 kilohms-I-A | | | | |
| | OXO.467.003TY | Resistor | 20 kilohms | | | |
| | | MJT-1-20,000-I-A | | | | |
| | | Resistor | 18 kilohms | | | |
| | | MJT-1-18 kilohms-I-A | | | | |
| | OXO.467.003TY | Resistor | 24 kilohms | | | |
| | | MJT-1-24,000-I-A | | | | |
| | | Resistor | 22 kilohms | | | |
| | | MJT-1-22 kilohms-I-A | | | | |
| | OXO.467.003TY | Resistor | 9.1 kilohms | | | |
| | | MJT-1-9.1 kilohms-I-A | | | | |
| 60 | OXO.467.003TY | Resistor | 62,000 ohms | 1 One of ratings | Same, 68,400 ohms $\pm 1\%$ | |
| | | MJT-1-62,000-I-A | | | | |
| | | Resistor | 5.1 kilohms | | | |
| | | MJT-1-5.1 kilohms-I-A | | | | |
| | OXO.467.003TY | Resistor | 3000 ohms | | | |
| | | MJT-1-3000-I | | | | |
| | | Resistor | 6.2 kilohms | | | |
| | | MJT-1-6.2 kilohms-I-A | | | | |
| | OXO.467.003TY | Resistor | 4300 ohms | | | |
| | | MJT-1-4300-I | | | | |
| | | Resistor | 6.8 kilohms | | | |
| | | MJT-1-6.8 kilohms-I-A | | | | |
| | OXO.467.003TY | Resistor | 5600 ohms | | | |
| | | MJT-1-5600-I | | | | |

21

SECRET

50X1-HUM

50X1-HUM

| 1 | 2 | 3 | 4 | 5 | 6 |
|----|---------------|-------------------------|--------------|---------------------------|--|
| | OXO.467.003TV | Resistor | 750 ohms | One of rat- ings | Selected in unit, series- connected, 10,000 ohms +1% |
| | | MNT-1-7500-I | | | |
| | OXO.467.003TV | Resistor | 1000 ohms | | |
| | | MNT-1-9100-I | | | |
| 61 | OXO.467.003TV | Resistor | 20 kilohms | 1 | Same, |
| | | MNT-1-20,000-I-A | | One | 10,000 ohms |
| | | Resistor | 1.2 kilohms | of | +1% |
| | | MNT-1-1.2 kilohms- | | rat- | |
| | | -I-A | | ings | |
| | OXO.467.003TV | Resistor | 620 ohms | | |
| | | MNT-1-620-I | | | |
| | | Resistor | 1.8 kilohms | | |
| | | MNT-1-1.8 kilohms- | | | |
| | | -I-A | | | |
| | OXO.467.003TV | Resistor | 1000 ohms | | |
| | | MNT-1-1000-I | | | |
| | OXO.467.003TV | Resistor | 1500 ohms | | |
| | | MNT-1-1500-I | | | |
| | OXO.467.003TV | Resistor | 2000 ohms | | |
| | | MNT-1-2000-I | | | |
| | OXO.467.003TV | Resistor | 2400 ohms | | |
| | | MNT-1-2400-I | | | |
| 62 | OXO.467.003TV | Resistor | 9100 ohms | 1 | Same, |
| | | MNT-1-9100-I | | One | 10,000 ohms |
| | OXO.467.003TV | Resistor | 430 ohms | of | +1% |
| | | MNT-1-430-I | | rat- | |
| | OXO.467.003TV | Resistor | 620 ohms | ings | |
| | | MNT-1-620-I | | | |
| | OXO.467.003TV | Resistor | 820 ohms | | |
| | | MNT-1-820-I | | | |
| | OXO.467.003TV | Resistor | 1000 ohms | | |
| | | MNT-1-1000-I | | | |
| | OXO.467.003TV | Resistor | 1200 ohms | | |
| | | MNT-1-1200-I | | | |
| 63 | OXO.467.003TV | Resistor | 5.1 kilohms | 1 | |
| | | MNT-1-5.1-I-A | | | |
| 64 | 4TV01-108-58 | Valve 6X2P | | 1 | |
| 65 | W6.732.008 | Wire-wound resistor | 5 ohms+1% | 1 | |
| 66 | FOCT 6118-52 | Capacitor | 0.01 μ F | 1 | |
| | | KBT-W-600-0.01-II | | | |
| 67 | OXO.468.004TV | Resistor | 10 kilohms | 1 | |
| | | CH-I-1-10A ₂ | | | |
| 68 | 4TV01-105-53 | Valve 6H1P | | 1 | |
| 69 | OXO.468.004TV | Resistor | 10 kilohms | 1 | |
| | | CH-I-1-10A-13 | | | |

22

SECRET

50X1-HUM

50X1-HUM

| 1 | 2 | 3 | 4 | 5 | 6 | 7 |
|----|-----------------------|-------------------------------------|--------------------------|---|--|---|
| 70 | ОЭО.467.003ТВ | Resistor МЛТ-1-3000-I | 3000 ohms | 1 | | |
| 71 | ОЭО.467.003ТВ | Resistor МЛТ-1-3000-I | 3000 ohms | 1 | | |
| 72 | ОЭО.467.003ТВ | Resistor МЛТ-1-75,000-I-A | 75 kilohms | 1 | | |
| 73 | ОЭО.467.003ТВ | Resistor МЛТ-1-2200-I | 2.2 kilohms | 1 | | |
| 74 | ОЭО.468.004ТВ | Resistor СН-1-1-10A ₄ | 10 kilohms | 1 | | |
| 75 | ГОСТ 5561-54 | Capacitor КЗ-2-20-100M | 100 μ F | 1 | | |
| 76 | НО.360.006 | Switch 5М6Н-КВ | 5 positions, 3 wafers | 1 | Porcelain | |
| 77 | ОЭО.467.003ТВ | Resistor МЛТ-1-1600-I | 1600 ohms | 1 | | |
| 78 | ГОСТ 5561-54 | Capacitor КЗ-16-8-500M | 500 μ F | 1 | | |
| 79 | ТУИНОПТ 533.077-54 | Micronmmeter, type M24 | 0-100 μ A | 1 | Inner resistance (1.900 ohms, cl. 1.5) | |
| 80 | ОЭО.468.004ТВ | Resistor СН-1-1-47A ₄ | 47 kilohms | 1 | | |
| 81 | ОЭО.467.003ТВ | Resistor МЛТ-1-2200-I | 2200 ohms | 1 | | |

SECRET

50X1-HUM

50X1-HUM

Page Denied

Next 2 Page(s) In Document Denied